

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW062322\
 Data File : VW026541.D
 Acq On : 24 Jun 2022 09:24
 Operator : SY/MD
 Sample : VW0623WBL03
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 61 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK385

Quant Time: Jun 25 02:31:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR062322WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Jun 25 02:21:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	204360	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	191849	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	82476	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.317	65	74008	4.503	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	90.000%
7) Chloroethane-d5	1.577	69	62367	4.724	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.400%
11) 1,1-Dichloroethene-d2	2.114	63	101291	3.468	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.400%
20) 2-Butanone-d5	3.928	46	83572	51.147	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	102.300%
24) Chloroform-d	4.352	84	125122	4.657	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.200%
26) 1,2-Dichloroethane-d4	5.037	65	50013	4.833	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.600%
32) Benzene-d6	5.053	84	251161	4.558	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.200%
36) 1,2-Dichloropropane-d6	6.072	67	74136	4.892	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.800%
41) Toluene-d8	7.314	98	209285	4.223	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.400%
43) trans-1,3-Dichloroprop...	7.625	79	20637	5.269	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	105.400%
46) 2-Hexanone-d5	8.092	63	64124	42.981	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	85.960%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	46153	4.656	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.200%
66) 1,2-Dichlorobenzene-d4	11.622	152	67638	4.956	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV062322\
Data File : VV026541.D
Acq On : 24 Jun 2022 09:24
Operator : SY/MD
Sample : VV0623WBL03
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 61 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK385

Quant Time: Jun 25 02:31:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR062322WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jun 25 02:21:54 2022
Response via : Initial Calibration

