

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW062322\
 Data File : VW026544.D
 Acq On : 24 Jun 2022 10:41
 Operator : SY/MD
 Sample : N3401-09
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AE1

Quant Time: Jun 25 02:32:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR062322WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Jun 25 02:21:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.625	114	173444	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	164027	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	66984	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.320	65	49944	3.581	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	71.600%	
7) Chloroethane-d5	1.581	69	47396	4.230	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	84.600%	
11) 1,1-Dichloroethene-d2	2.121	63	75195	3.034	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	60.600%	
20) 2-Butanone-d5	3.908	46	90084	64.960	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	129.920%	
24) Chloroform-d	4.362	84	113840	4.993	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	99.800%	
26) 1,2-Dichloroethane-d4	5.043	65	48110	5.478	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	109.600%	
32) Benzene-d6	5.060	84	215610	4.576	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	91.600%	
36) 1,2-Dichloropropane-d6	6.076	67	69110	5.334	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	106.600%	
41) Toluene-d8	7.320	98	171610	4.050	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	81.000%	
43) trans-1,3-Dichloroprop...	7.629	79	13958	4.168	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	83.400%	
46) 2-Hexanone-d5	8.092	63	57616	45.169	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	90.340%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	44055	5.198	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	104.000%	
66) 1,2-Dichlorobenzene-d4	11.622	152	57050	5.147	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	103.000%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV062322\
Data File : VV026544.D
Acq On : 24 Jun 2022 10:41
Operator : SY/MD
Sample : N3401-09
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AE1

Quant Time: Jun 25 02:32:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR062322WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jun 25 02:21:54 2022
Response via : Initial Calibration

