

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW062322\
 Data File : VW026545.D
 Acq On : 24 Jun 2022 11:05
 Operator : SY/MD
 Sample : N3401-11
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AE3

Quant Time: Jun 25 02:33:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR062322WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Jun 25 02:21:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.622	114	218286	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	204567	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	84270	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.320	65	80053	4.560	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	91.200%	
7) Chloroethane-d5	1.581	69	69905	4.957	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	99.200%	
11) 1,1-Dichloroethene-d2	2.121	63	107986	3.462	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	69.200%	
20) 2-Butanone-d5	3.909	46	103441	59.268	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	118.540%	
24) Chloroform-d	4.359	84	139129	4.848	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	97.000%	
26) 1,2-Dichloroethane-d4	5.044	65	58404	5.284	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	105.600%	
32) Benzene-d6	5.060	84	277395	4.721	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	94.400%	
36) 1,2-Dichloropropane-d6	6.076	67	81790	5.062	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	101.200%	
41) Toluene-d8	7.320	98	229526	4.344	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	86.800%	
43) trans-1,3-Dichloroprop...	7.629	79	19964	4.780	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	95.600%	
46) 2-Hexanone-d5	8.092	63	70557	44.352	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	88.700%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	51035	4.828	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	96.600%	
66) 1,2-Dichlorobenzene-d4	11.622	152	72591	5.205	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	104.200%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV062322\
Data File : VV026545.D
Acq On : 24 Jun 2022 11:05
Operator : SY/MD
Sample : N3401-11
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AE3

Quant Time: Jun 25 02:33:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR062322WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jun 25 02:21:54 2022
Response via : Initial Calibration

