

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV062322\
 Data File : VV026549.D
 Acq On : 24 Jun 2022 12:41
 Operator : SY/MD
 Sample : VIBLK260
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK260

Quant Time: Jun 25 06:26:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR062322WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Jun 25 02:21:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.622	114	250355	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	221637	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	95781	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.317	65	90122	4.476	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	89.600%	
7) Chloroethane-d5	1.577	69	74057	4.579	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	91.600%	
11) 1,1-Dichloroethene-d2	2.118	63	120611	3.371	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	67.400%	
20) 2-Butanone-d5	3.902	46	118543	59.221	ug/L	-0.01
Spiked Amount	50.000	Range 40 - 130	Recovery	=	118.440%	
24) Chloroform-d	4.359	84	151323	4.598	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	92.000%	
26) 1,2-Dichloroethane-d4	5.040	65	62846	4.958	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	99.200%	
32) Benzene-d6	5.056	84	323297	5.078	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	101.600%	
36) 1,2-Dichloropropane-d6	6.076	67	92056	5.258	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	105.200%	
41) Toluene-d8	7.320	98	268110	4.683	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	93.600%	
43) trans-1,3-Dichloroprop...	7.629	79	21871	4.833	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	96.600%	
46) 2-Hexanone-d5	8.092	63	89313	51.818	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	103.640%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	55845	4.876	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	97.600%	
66) 1,2-Dichlorobenzene-d4	11.622	152	83910	5.294	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	105.800%	
Target Compounds						
16) Methylene chloride	2.519	84	8791	0.490	ug/L	95
35) Methylcyclohexane	6.137	83	2720	0.108	ug/L #	91
42) Toluene	7.397	91	16436	0.246	ug/L	96
52) Ethylbenzene	9.024	91	2800	0.041	ug/L	85
53) m,p-Xylene	9.149	106	1602	0.059	ug/L	80
60) Isopropylbenzene	9.931	105	15325	0.259	ug/L	98
63) 1,2,4-Trimethylbenzene	10.918	105	3875	0.084	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV062322\
Data File : VV026549.D
Acq On : 24 Jun 2022 12:41
Operator : SY/MD
Sample : VIBLK260
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK260

Quant Time: Jun 25 06:26:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR062322WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jun 25 02:21:54 2022
Response via : Initial Calibration

