

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW062322\
 Data File : VW026551.D
 Acq On : 24 Jun 2022 13:29
 Operator : SY/MD
 Sample : N3425-13
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COAH5

Quant Time: Jun 25 02:34:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR062322WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Jun 25 02:21:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.622	114	249719	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	224819	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	96991	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.320	65	80638	4.015	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	80.400%
7) Chloroethane-d5	1.581	69	70048	4.342	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	86.800%
11) 1,1-Dichloroethene-d2	2.121	63	109471	3.068	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	61.400%
20) 2-Butanone-d5	3.905	46	113173	56.682	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	113.360%
24) Chloroform-d	4.362	84	146800	4.472	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.400%
26) 1,2-Dichloroethane-d4	5.040	65	57548	4.551	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.000%
32) Benzene-d6	5.060	84	293885	4.551	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.000%
36) 1,2-Dichloropropane-d6	6.076	67	85182	4.797	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.000%
41) Toluene-d8	7.320	98	245477	4.227	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.600%
43) trans-1,3-Dichloroprop...	7.629	79	21312	4.643	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	92.800%
46) 2-Hexanone-d5	8.092	63	83857	47.964	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	95.920%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	53165	4.577	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	91.600%
66) 1,2-Dichlorobenzene-d4	11.622	152	77009	4.798	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.000%
Target Compounds						Qvalue
13) Acetone	2.198	43	8577	6.771	ug/L	90
17) Methyl tert-butyl Ether	2.786	73	4034	0.147	ug/L #	67
22) cis-1,2-Dichloroethene	3.928	96	6403	0.375	ug/L	99
23) Bromochloromethane	4.265	128	2460	0.350	ug/L	90
25) Chloroform	4.388	83	92119	2.787	ug/L	94
33) Benzene	5.114	78	3893	0.059	ug/L	100
34) Trichloroethene	5.925	95	8376	0.485	ug/L	96
38) Bromodichloromethane	6.516	83	91820	4.916	ug/L	99
42) Toluene	7.394	91	44196	0.652	ug/L	97
47) Tetrachloroethene	7.979	164	100465	7.312	ug/L	98
49) Dibromochloromethane	8.246	129	117140	11.595	ug/L	98
53) m,p-Xylene	9.146	106	1592	0.058	ug/L	77
59) Bromoform	9.731	173	58033	16.237	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV062322\
Data File : VV026551.D
Acq On : 24 Jun 2022 13:29
Operator : SY/MD
Sample : N3425-13
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AH5

Quant Time: Jun 25 02:34:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR062322WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jun 25 02:21:54 2022
Response via : Initial Calibration

