

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV062322\
 Data File : VV026563.D
 Acq On : 24 Jun 2022 18:11
 Operator : SY/MD
 Sample : N3275-16
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Jun 25 02:37:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR062322WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Jun 25 02:21:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.625	114	209441	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	193677	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	82007	5.000	ug/L	0.00

System Monitoring Compounds						
4) Vinyl Chloride-d3	1.320	65	63674	3.780	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	75.600%
7) Chloroethane-d5	1.584	69	58225	4.303	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	86.000%
11) 1,1-Dichloroethene-d2	2.121	63	90718	3.031	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	60.600%
20) 2-Butanone-d5	3.912	46	103897	62.043	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	124.080%
24) Chloroform-d	4.365	84	123868	4.499	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.000%
26) 1,2-Dichloroethane-d4	5.047	65	51954	4.899	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.000%
32) Benzene-d6	5.063	84	241928	4.349	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.000%
36) 1,2-Dichloropropane-d6	6.076	67	74729	4.885	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.600%
41) Toluene-d8	7.320	98	194226	3.882	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	77.600%
43) trans-1,3-Dichloroprop...	7.632	79	17884	4.523	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.400%
46) 2-Hexanone-d5	8.092	63	71376	47.390	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	94.780%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	47752	4.772	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.400%
66) 1,2-Dichlorobenzene-d4	11.625	152	64588	4.759	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	95.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV062322\
Data File : VV026563.D
Acq On : 24 Jun 2022 18:11
Operator : SY/MD
Sample : N3275-16
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Jun 25 02:37:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR062322WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jun 25 02:21:54 2022
Response via : Initial Calibration

