

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VV062322\  
 Data File : VV026537.D  
 Acq On : 24 Jun 2022 01:06  
 Operator : SY/MD  
 Sample : VV0623WBL02  
 Misc : 25.0mL/MSVOA\_V/WATER  
 ALS Vial : 38 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 VBLK384

Quant Time: Jun 24 06:27:27 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR062322WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Fri Jun 24 00:51:15 2022  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon  | Response | Conc     | Units | Dev(Min) |
|-------------------------------|--------|-------|----------|----------|-------|----------|
| -----                         |        |       |          |          |       |          |
| Internal Standards            |        |       |          |          |       |          |
| 1) 1,4-Difluorobenzene        | 5.622  | 114   | 226607   | 5.000    | ug/L  | 0.00     |
| 28) Chlorobenzene-d5          | 8.853  | 117   | 213659   | 5.000    | ug/L  | 0.00     |
| 58) 1,4-Dichlorobenzene-d4    | 11.249 | 152   | 90717    | 5.000    | ug/L  | 0.00     |
| System Monitoring Compounds   |        |       |          |          |       |          |
| 4) Vinyl Chloride-d3          | 1.317  | 65    | 79045    | 4.337    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 40 - 130 | Recovery | =     | 86.800%  |
| 7) Chloroethane-d5            | 1.580  | 69    | 66615    | 4.550    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 65 - 130 | Recovery | =     | 91.000%  |
| 11) 1,1-Dichloroethene-d2     | 2.117  | 63    | 111027   | 3.429    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 60 - 125 | Recovery | =     | 68.600%  |
| 20) 2-Butanone-d5             | 3.934  | 46    | 97777    | 53.966   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000 | Range | 40 - 130 | Recovery | =     | 107.940% |
| 24) Chloroform-d              | 4.358  | 84    | 137088   | 4.602    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 125 | Recovery | =     | 92.000%  |
| 26) 1,2-Dichloroethane-d4     | 5.043  | 65    | 56345    | 4.911    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 130 | Recovery | =     | 98.200%  |
| 32) Benzene-d6                | 5.059  | 84    | 278409   | 4.537    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 125 | Recovery | =     | 90.800%  |
| 36) 1,2-Dichloropropane-d6    | 6.075  | 67    | 82370    | 4.881    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 60 - 140 | Recovery | =     | 97.600%  |
| 41) Toluene-d8                | 7.320  | 98    | 232802   | 4.218    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 130 | Recovery | =     | 84.400%  |
| 43) trans-1,3-Dichloroprop... | 7.632  | 79    | 16734    | 3.836    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 55 - 130 | Recovery | =     | 76.800%  |
| 46) 2-Hexanone-d5             | 8.091  | 63    | 80770    | 48.612   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000 | Range | 45 - 130 | Recovery | =     | 97.220%  |
| 56) 1,1,2,2-Tetrachloroeth... | 10.217 | 84    | 53478    | 4.844    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 65 - 120 | Recovery | =     | 96.800%  |
| 66) 1,2-Dichlorobenzene-d4    | 11.622 | 152   | 74830    | 4.984    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 80 - 120 | Recovery | =     | 99.600%  |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VV062322\  
Data File : VV026537.D  
Acq On : 24 Jun 2022 01:06  
Operator : SY/MD  
Sample : VV0623WBL02  
Misc : 25.0mL/MSVOA\_V/WATER  
ALS Vial : 38 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
VBLK384

Quant Time: Jun 24 06:27:27 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR062322WMA.M  
Quant Title : TRACE VOA SFAM1.0  
QLast Update : Fri Jun 24 00:51:15 2022  
Response via : Initial Calibration

