

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV062325\
 Data File : VV038819.D
 Acq On : 23 Jun 2025 09:23
 Operator : SY/MD
 Sample : VV0623WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK229

Quant Time: Jun 24 01:40:11 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM061625.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Jun 21 03:37:28 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	6610	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.785	117	6099	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	2691	0.500	ug/L	0.00

System Monitoring Compounds						
2) Vinyl Chloride-d3	1.294	65	2995	0.477	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	96.000%
4) 1,2-Dichloroethane-d4	4.969	65	1515	0.492	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	98.000%
7) 1,2-Dichloropropane-d6	5.995	67	1841	0.533	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	106.000%
8) Toluene-d8	7.253	98	3573	0.512	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	102.000%
10) 1,1,2,2-Tetrachloroeth...	10.146	84	1237	0.510	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	102.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV062325\
Data File : VV038819.D
Acq On : 23 Jun 2025 09:23
Operator : SY/MD
Sample : VV0623WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK229

Quant Time: Jun 24 01:40:11 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM061625.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Jun 21 03:37:28 2025
Response via : Initial Calibration

