

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV062325\
 Data File : VV038820.D
 Acq On : 23 Jun 2025 10:00
 Operator : SY/MD
 Sample : Q2377-02
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 PW-B6-L66-061925-SIM

Quant Time: Jun 24 01:40:22 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM061625.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Jun 21 03:37:28 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	7436	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.785	117	6823	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	2936	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.294	65	2985	0.423	ug/L	0.00
Spiked Amount 0.500	Range 40 - 130		Recovery =	84.000%		
4) 1,2-Dichloroethane-d4	4.969	65	1476	0.426	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery =	86.000%		
7) 1,2-Dichloropropane-d6	5.995	67	1812	0.469	ug/L	0.00
Spiked Amount 0.500	Range 60 - 140		Recovery =	94.000%		
8) Toluene-d8	7.253	98	3588	0.460	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery =	92.000%		
10) 1,1,2,2-Tetrachloroeth...	10.146	84	1147	0.422	ug/L	0.00
Spiked Amount 0.500	Range 65 - 120		Recovery =	84.000%		

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV062325\
Data File : VV038820.D
Acq On : 23 Jun 2025 10:00
Operator : SY/MD
Sample : Q2377-02
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
PW-B6-L66-061925-SIM

Quant Time: Jun 24 01:40:22 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM061625.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Jun 21 03:37:28 2025
Response via : Initial Calibration

