

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV062419\
 Data File : VV011737.D
 Acq On : 24 Jun 2019 18:26
 Operator : SY/MD
 Sample : K3381-12DL 250X
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFCM6DL

Quant Time: Jun 25 09:04:50 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR062119WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Jun 25 08:19:08 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	154220	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	142577	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	60355	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	49754	4.86	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	97.20%
7) Chloroethane-d5	1.57	69	39122	5.44	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	108.80%
11) 1,1-Dichloroethene-d2	2.12	63	88051	4.23	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	84.60%
20) 2-Butanone-d5	3.94	46	158150	58.65	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	117.30%
24) Chloroform-d	4.40	84	102724	5.50	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	110.00%
26) 1,2-Dichloroethane-d4	5.08	65	61670	5.50	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	110.00%
32) Benzene-d6	5.09	84	210002	5.28	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.60%
36) 1,2-Dichloropropane-d6	6.12	67	66345	5.42	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	108.40%
41) Toluene-d8	7.36	98	178790	4.97	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.40%
43) trans-1,3-Dichloropropene-	7.66	79	21086	4.93	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	98.60%
46) 2-Hexanone-d5	8.13	63	120473	59.98	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	119.96%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	46623	5.47	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	109.40%
64) 1,2-Dichlorobenzene-d4	11.67	152	61649	5.41	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	108.20%

Target Compounds

					Ovalue
5) Vinyl chloride	1.32	62	6007	0.421 ug/L	84
12) 1,1-Dichloroethene	2.13	96	21567	2.260 ug/L #	54
19) 1,1-Dichloroethane	3.22	63	52117	2.337 ug/L	97
22) cis-1,2-Dichloroethene	3.96	96	141104	11.443 ug/L	99
29) 1,1,1-Trichloroethane	4.65	97	17579	0.958 ug/L	99
34) Trichloroethene	5.96	95	6516	0.520 ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV062419\
Data File : VV011737.D
Acq On : 24 Jun 2019 18:26
Operator : SY/MD
Sample : K3381-12DL 250X
Misc : 25ML/MSVOA V/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFCM6DL

Quant Time: Jun 25 09:04:50 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR062119WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jun 25 08:19:08 2019
Response via : Initial Calibration

