

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV062420\
 Data File : VV017098.D
 Acq On : 25 Jun 2020 02:41
 Operator : SY/MD
 Sample : L3102-02
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F9L51

Quant Time: Jun 25 05:54:16 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR061720WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Jun 25 05:27:24 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	167554	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	161589	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	74675	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	36611	3.64	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	72.80%
7) Chloroethane-d5	1.58	69	31259	4.28	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	85.60%
11) 1,1-Dichloroethene-d2	2.13	63	59896	3.13	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	62.60%
20) 2-Butanone-d5	3.93	46	115272	44.98	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	89.96%
24) Chloroform-d	4.38	84	97344	4.52	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.40%
26) 1,2-Dichloroethane-d4	5.06	65	52742	4.63	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.60%
32) Benzene-d6	5.08	84	182706	4.36	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.20%
36) 1,2-Dichloropropane-d6	6.10	67	55987	4.46	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	89.20%
41) Toluene-d8	7.34	98	163710	4.16	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.20%
45) trans-1,3-Dichloropropene-	7.64	79	21445	4.10	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	82.00%
48) 2-Hexanone-d5	8.11	63	81438	40.89	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	81.78%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	40671	4.88	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.60%
65) 1,2-Dichlorobenzene-d4	11.65	152	63973	4.63	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	92.60%

Target Compounds					Ovalue
13) Acetone	2.21	43	23500	18.138 ug/L	98
14) Carbon disulfide	2.32	76	3585	0.149 ug/L #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV062420\
Data File : VV017098.D
Acq On : 25 Jun 2020 02:41
Operator : SY/MD
Sample : L3102-02
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F9L51

Quant Time: Jun 25 05:54:16 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR061720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Jun 25 05:27:24 2020
Response via : Initial Calibration

