

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV062421\
 Data File : VV021560.D
 Acq On : 24 Jun 2021 10:16
 Operator : SY/MD
 Sample : VV0624MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK196

Quant Time: Jun 25 03:01:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM052721WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jun 25 03:01:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	468949	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	428235	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	217659	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	131443	42.573	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	85.140%
7) Chloroethane-d5	1.568	69	113537	49.434	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	98.860%
11) 1,1-Dichloroethene-d2	2.108	63	190306	36.072	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.140%
21) 2-Butanone-d5	3.899	46	234615	105.970	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	105.970%
24) Chloroform-d	4.349	84	285171	46.399	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.800%
26) 1,2-Dichloroethane-d4	5.031	65	178849	50.521	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.040%
32) Benzene-d6	5.050	84	589559	49.212	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.420%
36) 1,2-Dichloropropane-d6	6.069	67	190547	51.505	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	103.000%
41) Toluene-d8	7.317	98	538384	47.171	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.340%
43) trans-1,3-Dichloroprop...	7.622	79	91559	46.245	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.480%
47) 2-Hexanone-d5	8.088	63	163756	96.396	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	96.400%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	248220	48.334	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	96.660%
66) 1,2-Dichlorobenzene-d4	11.625	152	209754	46.359	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.720%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV062421\
Data File : VV021560.D
Acq On : 24 Jun 2021 10:16
Operator : SY/MD
Sample : VV0624MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK196

Quant Time: Jun 25 03:01:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM052721WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jun 25 03:01:05 2021
Response via : Initial Calibration

