

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV062518\
 Data File : VV006434.D
 Acq On : 25 Jun 2018 17:34
 Operator : SY/MD
 Sample : J3673-03 40X
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DAZ98

Quant Time: Jun 26 07:06:33 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR062518WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Jun 26 06:50:47 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	339307	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	306127	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	136623	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	95339	4.27	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	85.40%
7) Chloroethane-d5	1.59	69	77668	4.32	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	86.40%
11) 1,1-Dichloroethene-d2	2.13	63	129398	3.27	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.40%
20) 2-Butanone-d5	3.97	46	191049	45.70	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	91.40%
24) Chloroform-d	4.41	84	159256	4.19	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.80%
26) 1,2-Dichloroethane-d4	5.09	65	79637	4.47	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.40%
32) Benzene-d6	5.10	84	357323	4.49	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.80%
36) 1,2-Dichloropropane-d6	6.12	67	106888	4.50	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	90.00%
41) Toluene-d8	7.36	98	325783	4.31	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.20%
43) trans-1,3-Dichloropropene-	7.67	79	34680	3.97	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	79.40%
46) 2-Hexanone-d5	8.14	63	144362	41.65	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	83.30%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	59436	4.08	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	81.60%
64) 1,2-Dichlorobenzene-d4	11.68	152	116132	4.57	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	91.40%

Target Compounds

					Ovalue
16) Methylene chloride	2.54	84	7284	0.30 ug/L	97
17) Methyl tert-butyl Ether	2.81	73	10298	0.21 ug/L #	88
22) cis-1,2-Dichloroethene	3.97	96	6178	0.25 ug/L	92
34) Trichloroethene	5.96	95	274210	10.60 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV062518\
Data File : VV006434.D
Acq On : 25 Jun 2018 17:34
Operator : SY/MD
Sample : J3673-03 40X
Misc : 25 mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DAZ98

Quant Time: Jun 26 07:06:33 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR062518WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jun 26 06:50:47 2018
Response via : Initial Calibration

