

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV062518\
 Data File : VV006438.D
 Acq On : 25 Jun 2018 19:22
 Operator : SY/MD
 Sample : J3673-07 40X
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DAZA2

Quant Time: Jun 26 07:13:13 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR062518WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Jun 26 06:50:47 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	322012	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	290956	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	127879	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	94162	4.45	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	89.00%
7) Chloroethane-d5	1.58	69	74665	4.38	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	87.60%
11) 1,1-Dichloroethene-d2	2.13	63	138880	3.70	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	74.00%
20) 2-Butanone-d5	3.98	46	194136	48.93	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	97.86%
24) Chloroform-d	4.41	84	158269	4.38	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.60%
26) 1,2-Dichloroethane-d4	5.09	65	79091	4.68	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.60%
32) Benzene-d6	5.11	84	350444	4.63	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.60%
36) 1,2-Dichloropropane-d6	6.12	67	104807	4.64	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.80%
41) Toluene-d8	7.37	98	319453	4.45	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.00%
43) trans-1,3-Dichloropropene-	7.67	79	34358	4.13	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	82.60%
46) 2-Hexanone-d5	8.14	63	148512	45.08	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	90.16%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	62552	4.51	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.20%
64) 1,2-Dichlorobenzene-d4	11.68	152	111663	4.70	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	94.00%

Target Compounds

					Ovalue
5) Vinyl chloride	1.32	62	24861	0.95	ug/L 94
12) 1,1-Dichloroethene	2.14	96	27122	1.39	ug/L # 28
22) cis-1,2-Dichloroethene	3.97	96	509352	21.73	ug/L 99
34) Trichloroethene	5.97	95	10410516	423.52	ug/L 97
47) Tetrachloroethene	8.02	164	22005	1.18	ug/L 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV062518\
Data File : VV006438.D
Acq On : 25 Jun 2018 19:22
Operator : SY/MD
Sample : J3673-07 40X
Misc : 25 mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DAZA2

Quant Time: Jun 26 07:13:13 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR062518WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jun 26 06:50:47 2018
Response via : Initial Calibration

