

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV062518\
 Data File : VV006454.D
 Acq On : 26 Jun 2018 03:27
 Operator : SY/MD
 Sample : J3673-12DL 1000X
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DAZA7DL

Quant Time: Jun 26 09:36:02 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR062518WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Jun 26 06:50:47 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	314627	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	284040	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	120353	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	91581	4.43	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	88.60%
7) Chloroethane-d5	1.58	69	76464	4.59	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	91.80%
11) 1,1-Dichloroethene-d2	2.13	63	123349	3.37	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.40%
20) 2-Butanone-d5	3.98	46	186369	48.08	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	96.16%
24) Chloroform-d	4.41	84	156008	4.42	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.40%
26) 1,2-Dichloroethane-d4	5.09	65	78886	4.78	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.60%
32) Benzene-d6	5.11	84	347967	4.71	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.20%
36) 1,2-Dichloropropane-d6	6.12	67	103937	4.72	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.40%
41) Toluene-d8	7.37	98	313421	4.47	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.40%
43) trans-1,3-Dichloropropene-	7.67	79	29587	3.65	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	73.00%
46) 2-Hexanone-d5	8.14	63	133500	41.51	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	83.02%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	57281	4.23	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	84.60%
64) 1,2-Dichlorobenzene-d4	11.68	152	107881	4.82	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.40%

Target Compounds					Ovalue
18) trans-1,2-Dichloroethene	2.79	96	16213	0.78 ug/L	98
22) cis-1,2-Dichloroethene	3.97	96	85840	3.75 ug/L	100
34) Trichloroethene	5.96	95	617946	25.75 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV062518\
Data File : VV006454.D
Acq On : 26 Jun 2018 03:27
Operator : SY/MD
Sample : J3673-12DL 1000X
Misc : 25 mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DAZA7DL

Quant Time: Jun 26 09:36:02 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR062518WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jun 26 06:50:47 2018
Response via : Initial Calibration

