

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV062519\
 Data File : VV011746.D
 Acq On : 25 Jun 2019 13:29
 Operator : SY/MD
 Sample : K3462-14DL 4X
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFCP9DL

Quant Time: Jun 26 01:52:14 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR062119WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Jun 26 01:40:21 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	163306	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	146084	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	59136	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	50381	4.65	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	93.00%
7) Chloroethane-d5	1.58	69	39895	5.24	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	104.80%
11) 1,1-Dichloroethene-d2	2.12	63	79169	3.59	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	71.80%
20) 2-Butanone-d5	3.96	46	156281	54.74	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	109.48%
24) Chloroform-d	4.40	84	114281	5.78	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	115.60%
26) 1,2-Dichloroethane-d4	5.08	65	63673	5.36	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.20%
32) Benzene-d6	5.09	84	230600	5.66	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	113.20%
36) 1,2-Dichloropropane-d6	6.12	67	72324	5.76	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	115.20%
41) Toluene-d8	7.36	98	194227	5.27	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.40%
43) trans-1,3-Dichloropropene-	7.66	79	21923	5.00	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	100.00%
46) 2-Hexanone-d5	8.14	63	115701	56.22	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	112.44%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	47457	5.44	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	108.80%
64) 1,2-Dichlorobenzene-d4	11.67	152	64348	5.77	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	115.40%

Target Compounds					Ovalue
34) Trichloroethene	5.95	95	62831	4.892 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV062519\
Data File : VV011746.D
Acq On : 25 Jun 2019 13:29
Operator : SY/MD
Sample : K3462-14DL 4X
Misc : 25ML/MSVOA V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFCP9DL

Quant Time: Jun 26 01:52:14 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR062119WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Jun 26 01:40:21 2019
Response via : Initial Calibration

