

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV062519\
 Data File : VV011747.D
 Acq On : 25 Jun 2019 13:57
 Operator : SY/MD
 Sample : K3462-01DL 10000X
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFCL0DL

Quant Time: Jun 26 01:58:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR062119WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Jun 26 01:40:21 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	161008	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	149503	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	62229	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	49359	4.62	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	92.40%
7) Chloroethane-d5	1.57	69	35686	4.75	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	95.00%
11) 1,1-Dichloroethene-d2	2.12	63	81845	3.77	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	75.40%
20) 2-Butanone-d5	3.95	46	152913	54.32	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	108.64%
24) Chloroform-d	4.40	84	105838	5.43	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	108.60%
26) 1,2-Dichloroethane-d4	5.08	65	63062	5.38	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.60%
32) Benzene-d6	5.09	84	218791	5.25	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.00%
36) 1,2-Dichloropropane-d6	6.12	67	68430	5.33	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	106.60%
41) Toluene-d8	7.36	98	185896	4.93	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.60%
43) trans-1,3-Dichloropropene-	7.66	79	20837	4.65	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	93.00%
46) 2-Hexanone-d5	8.14	63	113336	53.81	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	107.62%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	47154	5.28	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	105.60%
64) 1,2-Dichlorobenzene-d4	11.67	152	64591	5.50	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	110.00%

Target Compounds

					Ovalue
12) 1,1-Dichloroethene	2.13	96	9960	1.000 ug/L	# 1
19) 1,1-Dichloroethane	3.22	63	14262	0.612 ug/L	94
22) cis-1,2-Dichloroethene	3.96	96	26536	2.061 ug/L	99
29) 1,1,1-Trichloroethane	4.65	97	154040	8.002 ug/L	99
34) Trichloroethene	5.96	95	5212	0.397 ug/L	95
42) Toluene	7.43	91	36935	0.713 ug/L	99
47) Tetrachloroethene	8.02	164	1512	0.154 ug/L	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV062519\
Data File : VV011747.D
Acq On : 25 Jun 2019 13:57
Operator : SY/MD
Sample : K3462-01DL 10000X
Misc : 25ML/MSVOA V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFCL0DL

Quant Time: Jun 26 01:58:11 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR062119WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Jun 26 01:40:21 2019
Response via : Initial Calibration

