

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV062520\
 Data File : VV017120.D
 Acq On : 25 Jun 2020 12:15
 Operator : SY/MD
 Sample : VV0625WBL01
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK54

Quant Time: Jun 26 01:57:57 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR061720WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jun 26 01:31:34 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	182222	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	176458	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	82718	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	55101	5.04	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	100.80%
7) Chloroethane-d5	1.58	69	44075	5.54	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	110.80%
11) 1,1-Dichloroethene-d2	2.12	63	80949	3.89	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	77.80%
20) 2-Butanone-d5	3.96	46	115508	41.44	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	82.88%
24) Chloroform-d	4.37	84	119802	5.12	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.40%
26) 1,2-Dichloroethane-d4	5.06	65	63066	5.09	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.80%
32) Benzene-d6	5.07	84	230489	5.04	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.80%
36) 1,2-Dichloropropane-d6	6.09	67	67175	4.90	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.00%
41) Toluene-d8	7.33	98	205667	4.79	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.80%
45) trans-1,3-Dichloropropene-	7.64	79	27391	4.80	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	96.00%
48) 2-Hexanone-d5	8.11	63	98089	45.10	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	90.20%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	46306	5.08	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	101.60%
65) 1,2-Dichlorobenzene-d4	11.65	152	79600	5.21	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV062520\
Data File : VV017120.D
Acq On : 25 Jun 2020 12:15
Operator : SY/MD
Sample : VV0625WBL01
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK54

Quant Time: Jun 26 01:57:57 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR061720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jun 26 01:31:34 2020
Response via : Initial Calibration

