

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV062520\
 Data File : VV017157.D
 Acq On : 26 Jun 2020 03:59
 Operator : SY/MD
 Sample : L3068-07
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F9F32

Quant Time: Jun 26 06:43:51 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR061720WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jun 26 05:12:03 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	168821	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	164689	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	74503	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	50424	4.98	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	99.60%
7) Chloroethane-d5	1.58	69	39872	5.41	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	108.20%
11) 1,1-Dichloroethene-d2	2.13	63	74034	3.84	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	76.80%
20) 2-Butanone-d5	3.93	46	113931	44.12	ug/L	-0.04
Spiked Amount	50.000	Range	40 - 130	Recovery	=	88.24%
24) Chloroform-d	4.38	84	106182	4.90	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.00%
26) 1,2-Dichloroethane-d4	5.06	65	57639	5.02	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.40%
32) Benzene-d6	5.08	84	198076	4.64	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.80%
36) 1,2-Dichloropropane-d6	6.10	67	58509	4.57	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	91.40%
41) Toluene-d8	7.34	98	172951	4.32	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.40%
45) trans-1,3-Dichloropropene-	7.64	79	22017	4.13	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	82.60%
48) 2-Hexanone-d5	8.11	63	83031	40.91	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	81.82%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	41114	4.84	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.80%
65) 1,2-Dichlorobenzene-d4	11.65	152	67984	4.94	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV062520\
Data File : VV017157.D
Acq On : 26 Jun 2020 03:59
Operator : SY/MD
Sample : L3068-07
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F9F32

Quant Time: Jun 26 06:43:51 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR061720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jun 26 05:12:03 2020
Response via : Initial Calibration

