

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV062520\
 Data File : VV017168.D
 Acq On : 26 Jun 2020 09:05
 Operator : SY/MD
 Sample : VV0625WBL03
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK56

Quant Time: Jun 26 13:08:31 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR061720WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jun 26 12:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	171855	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	169909	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	79573	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	49237	4.77	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	95.40%
7) Chloroethane-d5	1.58	69	39250	5.23	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	104.60%
11) 1,1-Dichloroethene-d2	2.12	63	71991	3.67	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	73.40%
20) 2-Butanone-d5	3.95	46	122168	46.47	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	92.94%
24) Chloroform-d	4.37	84	110336	5.00	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.00%
26) 1,2-Dichloroethane-d4	5.06	65	58110	4.97	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.40%
32) Benzene-d6	5.07	84	201595	4.57	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.40%
36) 1,2-Dichloropropane-d6	6.09	67	60796	4.61	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.20%
41) Toluene-d8	7.33	98	174507	4.22	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.40%
45) trans-1,3-Dichloropropene-	7.64	79	23417	4.26	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.20%
48) 2-Hexanone-d5	8.11	63	85686	40.92	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	81.84%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	40476	4.62	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.40%
65) 1,2-Dichlorobenzene-d4	11.65	152	69575	4.73	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	94.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV062520\
Data File : VV017168.D
Acq On : 26 Jun 2020 09:05
Operator : SY/MD
Sample : VV0625WBL03
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 53 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK56

Quant Time: Jun 26 13:08:31 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR061720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jun 26 12:57:45 2020
Response via : Initial Calibration

