

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV062618\
 Data File : VV006464.D
 Acq On : 26 Jun 2018 13:53
 Operator : SY/MD
 Sample : J3673-19
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DAZB2

Quant Time: Jun 27 02:32:32 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR062518WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Jun 27 02:22:39 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	313283	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	283288	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	125613	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	88614	4.30	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	86.00%
7) Chloroethane-d5	1.58	69	69088	4.16	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	83.20%
11) 1,1-Dichloroethene-d2	2.13	63	120959	3.32	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	66.40%
20) 2-Butanone-d5	3.96	46	192110	49.77	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	99.54%
24) Chloroform-d	4.41	84	158490	4.51	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.20%
26) 1,2-Dichloroethane-d4	5.09	65	78752	4.79	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.80%
32) Benzene-d6	5.11	84	346981	4.71	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.20%
36) 1,2-Dichloropropane-d6	6.12	67	103585	4.71	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.20%
41) Toluene-d8	7.37	98	315325	4.51	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.20%
43) trans-1,3-Dichloropropene-	7.67	79	30557	3.78	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	75.60%
46) 2-Hexanone-d5	8.14	63	145590	45.39	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	90.78%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	59382	4.40	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	88.00%
64) 1,2-Dichlorobenzene-d4	11.68	152	114548	4.91	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.20%

Target Compounds					Ovalue
13) Acetone	2.21	43	66848	28.25 ug/L	98
21) 2-Butanone	4.05	43	20408	4.77 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV062618\
Data File : VV006464.D
Acq On : 26 Jun 2018 13:53
Operator : SY/MD
Sample : J3673-19
Misc : 25 mL/MSVOA V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DAZB2

Quant Time: Jun 27 02:32:32 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR062518WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Jun 27 02:22:39 2018
Response via : Initial Calibration

