

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV062618\
 Data File : VV006459.D
 Acq On : 26 Jun 2018 10:49
 Operator : SY/MD
 Sample : VV0626WBL01
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK64

Quant Time: Jun 27 02:27:05 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR062518WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Jun 27 02:22:39 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	320268	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	287440	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	123072	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	93124	4.42	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	88.40%
7) Chloroethane-d5	1.58	69	78837	4.65	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.00%
11) 1,1-Dichloroethene-d2	2.13	63	126910	3.40	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	68.00%
20) 2-Butanone-d5	3.96	46	144895	36.72	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	73.44%
24) Chloroform-d	4.41	84	158501	4.41	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.20%
26) 1,2-Dichloroethane-d4	5.09	65	77651	4.62	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.40%
32) Benzene-d6	5.10	84	357976	4.79	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.80%
36) 1,2-Dichloropropane-d6	6.12	67	106109	4.76	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.20%
41) Toluene-d8	7.36	98	322102	4.54	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.80%
43) trans-1,3-Dichloropropene-	7.67	79	32295	3.93	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	78.60%
46) 2-Hexanone-d5	8.14	63	134587	41.36	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	82.72%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	56694	4.14	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	82.80%
64) 1,2-Dichlorobenzene-d4	11.68	152	113274	4.95	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.00%


Target Compounds					Ovalue
16) Methylene chloride	2.54	84	16385	0.72 ug/L	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV062618\
Data File : VV006459.D
Acq On : 26 Jun 2018 10:49
Operator : SY/MD
Sample : VV0626WBL01
Misc : 25 mL/MSVOA V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK64

Quant Time: Jun 27 02:27:05 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR062518WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Jun 27 02:22:39 2018
Response via : Initial Calibration





#16
Methylene chloride
Concen: 0.72 ug/L
RT: 2.54 min Scan# 514
Delta R.T. 0.00 min
Lab File: VV006459.D
Acq: 26 Jun 2018 10:49

Instrument :
MSVOA_V
ClientSampleId :
VBLK64

Tot Ion: 84 Resp: 16385

Ion	Ratio	Lower	Upper
84	100		
86	62.8	44.2	82.2
49	108.4	79.0	146.8

