

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV062619\
 Data File : VV011754.D
 Acq On : 26 Jun 2019 11:11
 Operator : SY/MD
 Sample : VV0626WBL01
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK44

Quant Time: Jun 26 12:19:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR062119WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Jun 26 12:11:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	181454	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	170906	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	71514	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	62131	5.16	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	103.20%
7) Chloroethane-d5	1.58	69	50585	5.97	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	119.40%
11) 1,1-Dichloroethene-d2	2.12	63	94646	3.87	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	77.40%
20) 2-Butanone-d5	3.96	46	178120	56.15	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	112.30%
24) Chloroform-d	4.40	84	118421	5.39	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	107.80%
26) 1,2-Dichloroethane-d4	5.08	65	69516	5.26	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.20%
32) Benzene-d6	5.09	84	250891	5.26	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.20%
36) 1,2-Dichloropropane-d6	6.11	67	76521	5.21	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.20%
41) Toluene-d8	7.36	98	216482	5.02	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.40%
43) trans-1,3-Dichloropropene-	7.66	79	24446	4.77	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	95.40%
46) 2-Hexanone-d5	8.13	63	136668	56.76	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	113.52%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	52941	5.18	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	103.60%
64) 1,2-Dichlorobenzene-d4	11.67	152	72419	5.37	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	107.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV062619\
Data File : VV011754.D
Acq On : 26 Jun 2019 11:11
Operator : SY/MD
Sample : VV0626WBL01
Misc : 25ML/MSVOA V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK44

Quant Time: Jun 26 12:19:01 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR062119WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Jun 26 12:11:41 2019
Response via : Initial Calibration

