

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV062620\
 Data File : VV017182.D
 Acq On : 26 Jun 2020 15:30
 Operator : SY/MD
 Sample : L3103-01
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F9Q02

Quant Time: Jun 27 00:40:21 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Jun 27 00:23:55 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	296785	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	288287	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	131864	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	103972	49.65	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	99.30%
7) Chloroethane-d5	1.58	69	70848	56.38	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	112.76%
11) 1,1-Dichloroethene-d2	2.12	63	160251	45.57	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	91.14%
21) 2-Butanone-d5	3.90	46	176518	113.88	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	113.88%
24) Chloroform-d	4.37	84	245484	57.65	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	115.30%
26) 1,2-Dichloroethane-d4	5.06	65	179367	61.93	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	123.86%
32) Benzene-d6	5.07	84	466671	56.16	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	112.32%
36) 1,2-Dichloropropane-d6	6.09	67	141489	54.62	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	109.24%
41) Toluene-d8	7.33	98	422769	55.44	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	110.88%
43) trans-1,3-Dichloropropene-	7.64	79	73880	54.76	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	109.52%
47) 2-Hexanone-d5	8.11	63	117748	104.26	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	104.26%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	184730	53.87	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	107.74%
64) 1,2-Dichlorobenzene-d4	11.65	152	172854	61.42	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	122.84%#

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV062620\
Data File : VV017182.D
Acq On : 26 Jun 2020 15:30
Operator : SY/MD
Sample : L3103-01
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F9Q02

Quant Time: Jun 27 00:40:21 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Jun 27 00:23:55 2020
Response via : Initial Calibration

