

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV062620\
 Data File : VV017197.D
 Acq On : 26 Jun 2020 21:18
 Operator : SY/MD
 Sample : L3103-16
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F9Q73

Quant Time: Jun 27 01:52:33 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Jun 27 00:23:55 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	348334	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	333396	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	151889	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	100653	40.95	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	81.90%
7) Chloroethane-d5	1.58	69	72501	49.16	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	98.32%
11) 1,1-Dichloroethene-d2	2.12	63	156008	37.80	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	75.60%
21) 2-Butanone-d5	3.91	46	165617	91.04	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	91.04%
24) Chloroform-d	4.38	84	228316	45.68	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.36%
26) 1,2-Dichloroethane-d4	5.06	65	167245	49.20	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.40%
32) Benzene-d6	5.08	84	430035	44.75	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.50%
36) 1,2-Dichloropropane-d6	6.09	67	130700	43.63	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	87.26%
41) Toluene-d8	7.34	98	395026	44.79	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.58%
43) trans-1,3-Dichloropropene-	7.64	79	68215	43.72	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.44%
47) 2-Hexanone-d5	8.11	63	109125	83.55	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	83.55%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	172434	43.48	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	86.96%
64) 1,2-Dichlorobenzene-d4	11.65	152	160561	49.53	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.06%

Target Compounds					Ovalue
20) cis-1,2-Dichloroethene	3.94	96	35543	13.831 ug/L	97
34) Trichloroethene	5.94	95	61512	23.251 ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV062620\
Data File : VV017197.D
Acq On : 26 Jun 2020 21:18
Operator : SY/MD
Sample : L3103-16
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F9Q73

Quant Time: Jun 27 01:52:33 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Jun 27 00:23:55 2020
Response via : Initial Calibration

