

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV062620\
 Data File : VV017219.D
 Acq On : 27 Jun 2020 06:58
 Operator : SY/MD
 Sample : L3133-07
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F9Q29

Quant Time: Jun 27 07:23:11 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Jun 27 03:31:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	347287	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	324704	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	152636	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	88445	36.09	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	72.18%
7) Chloroethane-d5	1.58	69	60468	41.12	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	82.24%
11) 1,1-Dichloroethene-d2	2.12	63	142377	34.60	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	69.20%
21) 2-Butanone-d5	3.90	46	148853	82.07	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	82.07%
24) Chloroform-d	4.38	84	216136	43.38	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.76%
26) 1,2-Dichloroethane-d4	5.06	65	159183	46.97	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.94%
32) Benzene-d6	5.08	84	404860	43.25	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.50%
36) 1,2-Dichloropropane-d6	6.09	67	122072	41.84	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	83.68%
41) Toluene-d8	7.34	98	368041	42.85	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	85.70%
43) trans-1,3-Dichloropropene-	7.64	79	60504	39.82	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.64%
47) 2-Hexanone-d5	8.11	63	96961	76.23	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	76.23%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	160753	41.62	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	83.24%
64) 1,2-Dichlorobenzene-d4	11.65	152	149366	45.85	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.70%

Target Compounds

					Ovalue
13) Acetone	2.19	43	23087	19.319 ug/L	98
20) cis-1,2-Dichloroethene	3.94	96	8547	3.336 ug/L	98
34) Trichloroethene	5.94	95	44805	17.389 ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV062620\
Data File : VV017219.D
Acq On : 27 Jun 2020 06:58
Operator : SY/MD
Sample : L3133-07
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F9Q29

Quant Time: Jun 27 07:23:11 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Jun 27 03:31:05 2020
Response via : Initial Calibration

