

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV062620\
 Data File : VV017186.D
 Acq On : 26 Jun 2020 17:03
 Operator : SY/MD
 Sample : L3103-05
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F9Q11

Quant Time: Jun 27 00:53:25 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Jun 27 00:23:55 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	310555	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	295889	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	140264	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	100418	45.82	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	91.64%
7) Chloroethane-d5	1.58	69	65901	50.12	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	100.24%
11) 1,1-Dichloroethene-d2	2.12	63	154078	41.87	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	83.74%
21) 2-Butanone-d5	3.91	46	150934	93.06	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	93.06%
24) Chloroform-d	4.38	84	222460	49.93	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.86%
26) 1,2-Dichloroethane-d4	5.06	65	159531	52.64	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.28%
32) Benzene-d6	5.08	84	419428	49.17	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.34%
36) 1,2-Dichloropropane-d6	6.09	67	127306	47.88	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.76%
41) Toluene-d8	7.34	98	380559	48.62	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.24%
43) trans-1,3-Dichloropropene-	7.64	79	64378	46.49	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.98%
47) 2-Hexanone-d5	8.11	63	99567	85.90	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	85.90%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	166527	47.31	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	94.62%
64) 1,2-Dichlorobenzene-d4	11.65	152	156699	52.35	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	104.70%

Target Compounds


					Ovalue
13) Acetone	2.19	43	13066	12.227 ug/L	93
20) cis-1,2-Dichloroethene	3.94	96	108516	47.363 ug/L	99
34) Trichloroethene	5.94	95	143117	60.955 ug/L	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV062620\
Data File : VV017186.D
Acq On : 26 Jun 2020 17:03
Operator : SY/MD
Sample : L3103-05
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F9Q11

Quant Time: Jun 27 00:53:25 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Jun 27 00:23:55 2020
Response via : Initial Calibration





#13

Acetone

Concen: 12.227 ug/L

RT: 2.19 min Scan# 343

Delta R.T. 0.00 min

Lab File: VV017186.D

Acq: 26 Jun 2020 17:03





Instrument :
MSVOA_V
ClientSampled :
F9Q11

Tot Ion: 43 Resp: 13066

Ion Ratio Lower Upper

43 100

58 31.1 0.0 55.2



#20

cis-1,2-Dichloroethene

Concen: 47.363 ug/L

RT: 3.94 min Scan# 886

Delta R.T. 0.00 min

Lab File: VV017186.D

Acq: 26 Jun 2020 17:03




Tgt Ion: 96 Resp: 108516




Ion Ratio Lower Upper

96 100

61 131.7 93.4 173.4

68 0.0 0.0 0.0





#34

Trichloroethene

Concen: 60.955 ug/L

RT: 5.94 min Scan# 1508

Delta R.T. 0.00 min

Lab File: VV017186.D

Acq: 26 Jun 2020 17:03

Instrument :

MSVOA_V

ClientSampleId :

F9Q11

Tot Ion: 95 Resp: 143117

Ion Ratio Lower Upper

95 100

97 63.2 44.3 82.3

132 95.0 67.9 126.1

130 102.1 72.2 134.2

