

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV062719\
 Data File : VV011764.D
 Acq On : 26 Jun 2019 17:16
 Operator : SY/MD
 Sample : VSTDICV0.5
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VICV09

Quant Time: Jun 27 05:11:00 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR062719SIM.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Jun 27 04:08:35 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	8636	0.50	ug/L	0.00
5) Chlorobenzene-d5	8.89	117	7655	0.50	ug/L	-0.01
10) 1,4-Dichlorobenzene-d4	11.29	152	3204	0.50	ug/L	-0.01

System Monitoring Compounds

2) Vinyl Chloride-d3	1.32	65	4371	0.49	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	98.00%
4) 1,2-Dichloroethane-d4	5.08	65	3705	0.51	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	102.00%
7) Toluene-d8	7.36	98	9581	0.50	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	100.00%
9) 1,1,2,2-Tetrachloroethane-	10.24	84	2636	0.52	ug/L	-0.01
Spiked Amount	0.500	Range	65 - 120	Recovery	=	104.00%
11) 1-2-dichlorobenzene-d4	11.67	152	3424	0.51	ug/L	-0.01
Spiked Amount	0.500	Range	80 - 120	Recovery	=	102.00%

Target Compounds

					Qvalue
3) Vinyl chloride	1.32	62	4377	0.501	ug/L 99
6) Trichloroethene	5.95	95	3465	0.509	ug/L 99
8) 1,2-Dibromoethane	8.39	107	2023	0.519	ug/L 98
12) 1,2,3-Trichloropropane	10.31	75	2139	0.534	ug/L 99
13) 1,2-Dibromo-3-chloropropan	12.47	75	356	0.519	ug/L 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV062719\
Data File : VV011764.D
Acq On : 26 Jun 2019 17:16
Operator : SY/MD
Sample : VSTDICV0.5
Misc : 25ML/MSVOA V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VICV09

Quant Time: Jun 27 05:11:00 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR062719SIM.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Jun 27 04:08:35 2019
Response via : Initial Calibration

