

Method Path : Z:\VOASRV\HPCHEM1\MSV0A V\METHOD\
 Method File : SOMVTR062719SIM.M
 Title : TRACE VOA SOM01.0
 Last Update : Thu Jun 27 04:08:35 2019
 Response Via : Initial Calibration

Calibration Files

0.5 =VV011761.D 0.05=VV011759.D 0.1 =VV011760.D
 1.0 =VV011762.D 2 =VV011763.D

	Compound	0.5	0.05	0.1	1.0	2	Avg	%RSD
1) I	1,4-Difluorobenzene	-----ISTD-----						
2) S	Vinyl Chloride-d3	0.484	0.513	0.557	0.516	0.517	0.517	5.03
3) T	Vinyl chloride	0.487	0.495	0.496	0.527	0.523	0.505	3.59
4) S	1,2-Dichloroethane-	0.410	0.387	0.432	0.428	0.435	0.418	4.82
5) I	Chlorobenzene-d5	-----ISTD-----						
6) T	Trichloroethene	0.442	0.433	0.423	0.462	0.461	0.444	3.90
7) S	Toluene-d8	1.206	1.141	1.273	1.261	1.317	1.240	5.47
8) T	1,2-Dibromoethane	0.251	0.236	0.247	0.266	0.272	0.254	5.82
9) S	1,1,2,2-Tetrachloro	0.333	0.291	0.340	0.343	0.357	0.333	7.55
10) I	1,4-Dichlorobenzene-d	-----ISTD-----						
11) S	1-2-dichlorobenzene	1.000	1.163	1.054	1.025	1.047	1.058	5.91
12)	1,2,3-Trichloroprop	0.637	0.567	0.601	0.655	0.668	0.626	6.61
13) T	1,2-Dibromo-3-chlor	0.102	0.109	0.106	0.109	0.109	0.107	2.99

(#) = Out of Range