

Data Path : Z:\voasrv\HPCHEM1\MSVOA V\Data\VV062719\
 Data File : VV011764.D
 Acq On : 26 Jun 2019 17:16
 Operator : SY/MD
 Sample : VSTDICV0.5
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VICV09

Quant Time: Jun 27 05:11:00 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR062719SIM.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Jun 27 04:08:35 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Difluorobenzene	1.000	1.000	0.0	98	0.00
2 S	Vinyl Chloride-d3	0.517	0.506	2.1	103	0.00
3 T	Vinyl chloride	0.505	0.507	-0.4	102	0.00
4 S	1,2-Dichloroethane-d4	0.418	0.429	-2.6	103	0.00
5 I	Chlorobenzene-d5	1.000	1.000	0.0	99	-0.01
6 T	Trichloroethene	0.444	0.453	-2.0	101	0.00
7 S	Toluene-d8	1.240	1.252	-1.0	103	0.00
8 T	1,2-Dibromoethane	0.254	0.264	-3.9	104	0.00
9 S	1,1,2,2-Tetrachloroethane-d	0.333	0.344	-3.3	103	-0.01
10 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	-0.01
11 S	1-2-dichlorobenzene-d4	1.058	1.069	-1.0	106	-0.01
12	1,2,3-Trichloropropane	0.626	0.668	-6.7	103	0.00
13 T	1,2-Dibromo-3-chloropropane	0.107	0.111	-3.7	108	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0