

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV062725\
 Data File : VV038829.D
 Acq On : 27 Jun 2025 12:11
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK

Quant Time: Jun 28 04:02:07 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V062725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 28 03:44:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.600	168	490365	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.532	114	898960	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.773	117	820520	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.172	152	380720	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.941	65	342899	54.318	ug/l	0.00
Spiked Amount	50.000		Recovery	=	108.640%	
35) Dibromofluoromethane	4.503	113	283637	48.630	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.260%	
50) Toluene-d8	7.236	98	872306	43.772	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.540%	
62) 4-Bromofluorobenzene	10.001	95	309488	42.187	ug/l	0.00
Spiked Amount	50.000		Recovery	=	84.380%	

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV062725\
Data File : VV038829.D
Acq On : 27 Jun 2025 12:11
Operator : SY/MD
Sample : VIBLK
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK

Quant Time: Jun 28 04:02:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V062725W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 28 03:44:11 2025
Response via : Initial Calibration

