

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV062818\
 Data File : VV006492.D
 Acq On : 28 Jun 2018 16:36
 Operator : SY/MD
 Sample : J3615-02
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BEXT8

Quant Time: Jun 29 04:05:24 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR062518WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jun 29 03:51:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	315105	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	285341	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	128014	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	113518	5.48	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	109.60%
7) Chloroethane-d5	1.58	69	90129	5.40	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	108.00%
11) 1,1-Dichloroethene-d2	2.13	63	146763	4.00	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	80.00%
20) 2-Butanone-d5	3.97	46	206739	53.25	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	106.50%
24) Chloroform-d	4.41	84	174276	4.93	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.60%
26) 1,2-Dichloroethane-d4	5.09	65	83949	5.08	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.60%
32) Benzene-d6	5.10	84	383925	5.17	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.40%
36) 1,2-Dichloropropane-d6	6.12	67	112182	5.07	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.40%
41) Toluene-d8	7.36	98	340998	4.84	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.80%
43) trans-1,3-Dichloropropene-	7.67	79	33127	4.06	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.20%
46) 2-Hexanone-d5	8.14	63	160211	49.59	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	99.18%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	71696	5.28	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	105.60%
64) 1,2-Dichlorobenzene-d4	11.68	152	122842	5.16	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.20%

Target Compounds					Ovalue
19) 1,1-Dichloroethane	3.23	63	10735	0.29 ug/L	97
33) Benzene	5.15	78	8375889	99.40 ug/L	100
42) Toluene	7.44	91	2065778	23.15 ug/L	100
51) Chlorobenzene	8.93	112	141534	2.40 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV062818\
Data File : VV006492.D
Acq On : 28 Jun 2018 16:36
Operator : SY/MD
Sample : J3615-02
Misc : 25 mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BEXT8

Quant Time: Jun 29 04:05:24 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR062518WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jun 29 03:51:20 2018
Response via : Initial Calibration

