

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV062918\
 Data File : VV006507.D
 Acq On : 29 Jun 2018 14:13
 Operator : SY/MD
 Sample : J3615-07
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Jun 30 02:38:20 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR062518WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Jun 30 02:15:06 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	330611	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	298104	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	129627	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	97846	4.50	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	90.00%
7) Chloroethane-d5	1.58	69	77571	4.43	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	88.60%
11) 1,1-Dichloroethene-d2	2.13	63	127318	3.31	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	66.20%
20) 2-Butanone-d5	3.96	46	194111	47.65	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	95.30%
24) Chloroform-d	4.41	84	158551	4.28	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.60%
26) 1,2-Dichloroethane-d4	5.09	65	78313	4.52	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.40%
32) Benzene-d6	5.11	84	349618	4.51	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.20%
36) 1,2-Dichloropropane-d6	6.12	67	104591	4.52	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	90.40%
41) Toluene-d8	7.37	98	309456	4.20	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.00%
43) trans-1,3-Dichloropropene-	7.67	79	30544	3.59	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	71.80%
46) 2-Hexanone-d5	8.14	63	139731	41.40	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	82.80%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	58179	4.10	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	82.00%
64) 1,2-Dichlorobenzene-d4	11.68	152	107981	4.48	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	89.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV062918\
Data File : VV006507.D
Acq On : 29 Jun 2018 14:13
Operator : SY/MD
Sample : J3615-07
Misc : 25 mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Jun 30 02:38:20 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR062518WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Jun 30 02:15:06 2018
Response via : Initial Calibration

