

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV062920\
 Data File : VV017228.D
 Acq On : 29 Jun 2020 10:13
 Operator : SY/MD
 Sample : L3103-20
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Jun 30 07:54:55 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Jun 30 07:40:27 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	360859	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	345624	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	165569	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	98231	38.58	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	77.16%
7) Chloroethane-d5	1.58	69	71512	46.80	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	93.60%
11) 1,1-Dichloroethene-d2	2.12	63	160234	37.47	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	74.94%
21) 2-Butanone-d5	3.91	46	161611	85.75	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	85.75%
24) Chloroform-d	4.38	84	240452	46.44	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.88%
26) 1,2-Dichloroethane-d4	5.05	65	171656	48.74	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.48%
32) Benzene-d6	5.07	84	451436	45.31	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.62%
36) 1,2-Dichloropropane-d6	6.09	67	135566	43.65	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	87.30%
41) Toluene-d8	7.34	98	413905	45.27	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.54%
43) trans-1,3-Dichloropropene-	7.64	79	74849	46.27	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.54%
47) 2-Hexanone-d5	8.11	63	109360	80.77	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	80.77%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	177825	43.25	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	86.50%
64) 1,2-Dichlorobenzene-d4	11.65	152	173908	49.22	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.44%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV062920\
Data File : VV017228.D
Acq On : 29 Jun 2020 10:13
Operator : SY/MD
Sample : L3103-20
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Jun 30 07:54:55 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Jun 30 07:40:27 2020
Response via : Initial Calibration

