

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV062920\
 Data File : VV017251.D
 Acq On : 30 Jun 2020 10:53
 Operator : SY/MD
 Sample : VV0629MBL01
 Misc : 5.0µ/5.0mL/100µL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK96

Quant Time: Jun 30 12:13:08 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Jun 30 12:05:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	352615	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	328405	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	158847	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	82234	33.05	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	66.10%
7) Chloroethane-d5	1.58	69	65434	43.83	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	87.66%
11) 1,1-Dichloroethene-d2	2.12	63	138664	33.19	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	66.38%
21) 2-Butanone-d5	3.91	46	164067	89.09	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	89.09%
24) Chloroform-d	4.37	84	222088	43.90	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.80%
26) 1,2-Dichloroethane-d4	5.05	65	160633	46.68	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.36%
32) Benzene-d6	5.07	84	404778	42.76	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.52%
36) 1,2-Dichloropropane-d6	6.09	67	125019	42.37	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	84.74%
41) Toluene-d8	7.33	98	365811	42.11	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	84.22%
43) trans-1,3-Dichloropropene-	7.64	79	67006	43.60	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.20%
47) 2-Hexanone-d5	8.11	63	107303	83.41	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	83.41%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	170877	43.74	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	87.48%
64) 1,2-Dichlorobenzene-d4	11.65	152	160892	47.46	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.92%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV062920\
Data File : VV017251.D
Acq On : 30 Jun 2020 10:53
Operator : SY/MD
Sample : VV0629MBL01
Misc : 5.0µL/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK96

Quant Time: Jun 30 12:13:08 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Jun 30 12:05:50 2020
Response via : Initial Calibration

