

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW062923\
 Data File : VW031614.D
 Acq On : 29 Jun 2023 11:49
 Operator : SY/MD
 Sample : 03421-11ME
 Misc : 3.51g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 A46G7ME

Quant Time: Jun 30 04:26:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM061423WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jun 30 04:25:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.539	114	214190	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.793	117	241635	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	116591	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	58518	41.051	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	82.100%
7) Chloroethane-d5	1.523	69	30090	24.192	ug/L	-0.01
Spiked Amount	50.000	Range	70 - 130	Recovery	=	48.380%#
11) 1,1-Dichloroethene-d2	2.050	65	32672	44.471	ug/L	-0.01
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.940%
21) 2-Butanone-d5	3.806	46	90626	78.514	ug/L	0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	78.510%
24) Chloroform-d	4.253	84	133377	37.360	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	74.720%
26) 1,2-Dichloroethane-d4	4.947	65	89524	40.079	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	80.160%
32) Benzene-d6	4.963	84	220217	32.378	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	64.760%#
36) 1,2-Dichloropropane-d6	5.995	67	83667	34.707	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	69.420%#
41) Toluene-d8	7.249	98	191128	30.622	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	61.240%#
43) trans-1,3-Dichloroprop...	7.561	79	42403	36.082	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.160%
47) 2-Hexanone-d5	8.050	63	70814	74.399	ug/L	0.02
Spiked Amount	100.000	Range	45 - 130	Recovery	=	74.400%
56) 1,1,2,2-Tetrachloroeth...	10.162	84	141680	38.937	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	77.880%
66) 1,2-Dichlorobenzene-d4	11.571	152	88945	40.022	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	80.040%
Target Compounds						
					Qvalue	
13) Acetone	2.130	43	5341	4.689	ug/L	92
15) Methyl Acetate	2.384	43	25925	13.519	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV062923\
Data File : VV031614.D
Acq On : 29 Jun 2023 11:49
Operator : SY/MD
Sample : 03421-11ME
Misc : 3.51g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
A46G7ME

Quant Time: Jun 30 04:26:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM061423WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jun 30 04:25:13 2023
Response via : Initial Calibration

