

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW063022\
 Data File : VW026604.D
 Acq On : 30 Jun 2022 13:55
 Operator : SY/MD
 Sample : N3521-10
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0JX2

Quant Time: Jul 01 00:22:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM062922WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jul 01 00:19:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.625	114	375703	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	373645	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	161754	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.320	65	133753	43.673	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	87.340%
7) Chloroethane-d5	1.584	69	122966	46.816	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	93.640%
11) 1,1-Dichloroethene-d2	2.121	63	198609	34.426	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	68.860%
21) 2-Butanone-d5	3.921	46	149257	97.341	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	97.340%
24) Chloroform-d	4.362	84	276259	46.716	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.440%
26) 1,2-Dichloroethane-d4	5.044	65	159829	47.499	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.000%
32) Benzene-d6	5.060	84	533378	46.945	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.880%
36) 1,2-Dichloropropane-d6	6.076	67	170175	47.552	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.100%
41) Toluene-d8	7.320	98	440632	44.153	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.300%
43) trans-1,3-Dichloroprop...	7.625	79	61483	42.791	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.580%
47) 2-Hexanone-d5	8.095	63	106027	85.663	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	85.660%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	236642	45.798	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.600%
66) 1,2-Dichlorobenzene-d4	11.622	152	170300	53.795	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	107.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV063022\
Data File : VV026604.D
Acq On : 30 Jun 2022 13:55
Operator : SY/MD
Sample : N3521-10
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0JX2

Quant Time: Jul 01 00:22:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM062922WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jul 01 00:19:36 2022
Response via : Initial Calibration

