

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW063022\
 Data File : VW026611.D
 Acq On : 30 Jun 2022 16:47
 Operator : SY/MD
 Sample : N3521-01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0JT3

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 07/01/2022
 Supervised By :Mahesh Dadoda 07/01/2022

Quant Time: Jul 01 00:23:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM062922WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jul 01 00:19:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.622	114	370168	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	368792	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	180040	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.320	65	119607	39.638	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	79.280%		
7) Chloroethane-d5	1.584	69	110702	42.777	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	85.560%		
11) 1,1-Dichloroethene-d2	2.121	63	178134	31.339	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	62.680%		
21) 2-Butanone-d5	3.915	46	154111	102.010	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	102.010%		
24) Chloroform-d	4.362	84	257944	44.271	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	88.540%		
26) 1,2-Dichloroethane-d4	5.040	65	153336	46.251	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	92.500%		
32) Benzene-d6	5.059	84	491284	43.809	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	87.620%		
36) 1,2-Dichloropropane-d6	6.076	67	161907	45.837	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	91.680%		
41) Toluene-d8	7.320	98	401094	40.720	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	81.440%		
43) trans-1,3-Dichloroprop...	7.625	79	56069	39.536	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	79.080%		
47) 2-Hexanone-d5	8.092	63	104025	85.152	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	85.150%		
56) 1,1,2,2-Tetrachloroeth...	10.214	84	235693	46.214	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	92.420%		
66) 1,2-Dichlorobenzene-d4	11.622	152	183994m	52.218	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	104.440%		
Target Compounds					Qvalue	
13) Acetone	2.208	43	11016	10.015	ug/L	97
33) Benzene	5.111	78	20407	1.707	ug/L	100
51) Chlorobenzene	8.882	112	374921	44.313	ug/L	99
64) 1,3-Dichlorobenzene	11.181	146	52732	9.496	ug/L	97
65) 1,4-Dichlorobenzene	11.271	146	210346	36.530	ug/L	99
67) 1,2-Dichlorobenzene	11.641	146	1647137	291.093	ug/L	100
70) 1,2,4-trichlorobenzene	13.259	180	296616	92.329	ug/L	99
72) 1,2,3-Trichlorobenzene	13.741	180	150117	46.028	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV063022\
Data File : VV026611.D
Acq On : 30 Jun 2022 16:47
Operator : SY/MD
Sample : N3521-01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0JT3

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 07/01/2022
Supervised By :Mahesh Dadoda 07/01/2022

Quant Time: Jul 01 00:23:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM062922WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jul 01 00:19:36 2022
Response via : Initial Calibration

