

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW063022\
 Data File : VW026610.D
 Acq On : 30 Jun 2022 16:21
 Operator : SY/MD
 Sample : N3534-07
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 A0AH4

Quant Time: Jul 01 00:23:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM062922WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jul 01 00:19:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.625	114	375299	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	366270	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	158202	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.320	65	122917	40.178	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	80.360%
7) Chloroethane-d5	1.584	69	113063	43.092	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	86.180%
11) 1,1-Dichloroethene-d2	2.121	63	181076	31.421	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	62.840%
21) 2-Butanone-d5	3.915	46	131034	85.549	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	85.550%
24) Chloroform-d	4.362	84	258530	43.765	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.540%
26) 1,2-Dichloroethane-d4	5.040	65	152236	45.292	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.580%
32) Benzene-d6	5.059	84	498118	44.724	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.440%
36) 1,2-Dichloropropane-d6	6.075	67	161113	45.927	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	91.860%
41) Toluene-d8	7.320	98	399019	40.789	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	81.580%
43) trans-1,3-Dichloroprop...	7.625	79	56566	40.161	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	80.320%
47) 2-Hexanone-d5	8.091	63	103167	85.031	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	85.030%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	230417	45.491	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	90.980%
66) 1,2-Dichlorobenzene-d4	11.622	152	157136	50.752	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.500%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV063022\
Data File : VV026610.D
Acq On : 30 Jun 2022 16:21
Operator : SY/MD
Sample : N3534-07
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
A0AH4

Quant Time: Jul 01 00:23:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM062922WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jul 01 00:19:36 2022
Response via : Initial Calibration

