


Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV063022\
 Data File : VV026597.D
 Acq On : 30 Jun 2022 09:22
 Operator : SY/MD
 Sample : BFB293
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFB293

Integration File: LSCINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM062922WMA.M
 Title : VOC Analysis
 Last Update : Fri Jul 01 00:19:36 2022



AutoFind: Scans 2809, 2810, 2811; Background Corrected with Scan 2799

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.9	6169	PASS
75	95	30	80	47.1	18272	PASS
95	95	100	100	100.0	38827	PASS
96	95	5	9	6.1	2354	PASS
173	174	0.00	2	1.7	589	PASS
174	95	50	120	87.2	33867	PASS
175	174	5	9	7.7	2617	PASS
176	174	95	105	98.0	33189	PASS
177	176	5	10	6.7	2209	PASS

m/z	Abundance
39.90	217.0
44.00	406.0

Instrument :
MSVOA_V
ClientSampleId :
BFB293

m/z	Abundance
36.00	265.0
37.00	1501.0
38.00	1385.0
39.10	570.0
39.80	267.0
44.00	617.0
44.90	276.0
47.10	534.0
47.90	288.0
49.00	1550.0
50.00	6559.0
51.00	1906.0
55.00	147.0
56.10	457.0
56.90	875.0
60.10	290.0
61.00	1851.0
62.00	1799.0
63.00	1486.0
64.10	314.0
65.00	317.0
66.90	119.0
68.00	3685.0
69.00	3998.0
69.90	331.0
71.90	183.0
73.00	1508.0
74.00	6020.0
75.10	18744.0
76.10	1735.0
76.90	260.0
77.90	182.0
78.90	1644.0
79.90	504.0
80.90	1719.0
82.00	369.0
87.00	2225.0
88.00	1772.0
90.90	200.0
91.90	960.0
93.10	1841.0
94.00	4874.0
95.10	39416.0
96.10	2640.0
103.90	131.0
106.00	215.0
115.90	201.0
116.90	358.0
117.80	147.0
118.80	222.0
127.90	152.0
129.90	166.0
141.00	361.0
142.90	654.0
145.00	267.0
148.00	142.0
171.00	117.0
172.00	564.0
173.00	635.0
174.00	33096.0
174.90	2635.0
176.00	33024.0
177.00	2187.0

Instrument :
MSVOA_V
ClientSampleId :
BFB293

m/z	Abundance
35.90	306.0
37.00	1508.0
38.00	1350.0
39.10	584.0
39.80	264.0
44.00	695.0
45.10	299.0
47.10	490.0
47.80	214.0
49.00	1581.0
50.00	6347.0
51.00	1961.0
51.90	128.0
55.00	152.0
56.10	522.0
56.90	805.0
60.00	335.0
61.00	1905.0
62.00	1732.0
63.00	1416.0
64.10	314.0
65.00	393.0
67.00	125.0
68.00	3685.0
69.00	3925.0
70.00	342.0
71.90	214.0
73.00	1514.0
74.00	6005.0
75.10	18920.0
76.10	1720.0
76.80	263.0
77.90	193.0
78.10	194.0
78.90	1622.0
79.90	534.0
80.90	1859.0
81.90	426.0
87.00	2132.0
88.00	1858.0
90.90	210.0
92.00	900.0
93.00	1894.0
94.00	4943.0
95.00	40296.0
96.10	2406.0
104.00	171.0
106.00	228.0
116.00	181.0
116.90	341.0
117.90	138.0
118.80	202.0
128.00	128.0
130.00	181.0
141.00	363.0
142.90	706.0
145.00	288.0
148.00	146.0
155.10	105.0
171.10	154.0
172.00	571.0
173.00	639.0
174.00	35120.0
175.00	2701.0
176.00	34384.0
177.00	2291.0

Instrument :
MSVOA_V
ClientSampleId :
BFB293

m/z	Abundance
35.90	346.0
37.00	1340.0
38.00	1155.0
39.00	561.0
39.90	282.0
44.00	707.0
45.10	298.0
46.00	129.0
47.10	461.0
47.80	168.0
49.00	1352.0
50.00	5601.0
51.00	1875.0
51.90	130.0
54.90	170.0
56.00	507.0
57.00	785.0
60.00	339.0
61.00	1803.0
62.00	1534.0
63.10	1399.0
63.90	258.0
64.90	417.0
67.00	109.0
68.00	3346.0
69.00	3315.0
70.00	305.0
71.90	206.0
73.00	1365.0
74.10	5614.0
75.00	17152.0
76.10	1520.0
76.80	277.0
77.90	177.0
78.90	1427.0
79.90	516.0
80.90	1746.0
81.90	453.0
87.00	1739.0
88.00	1749.0
90.80	196.0
92.10	852.0
93.00	1741.0
94.00	4418.0
95.00	36768.0
96.00	2015.0
102.90	102.0
104.00	200.0
106.00	202.0
115.90	171.0
116.80	315.0
118.00	134.0
118.90	216.0
127.80	128.0
130.00	171.0
141.00	347.0
142.90	696.0
145.10	267.0
147.80	116.0
154.70	107.0
171.10	152.0
172.00	559.0
173.00	492.0
174.00	33384.0
175.00	2515.0
176.00	32160.0
177.00	2148.0

Instrument :
MSVOA_V
ClientSampleId :
BFB293