

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV070120\
 Data File : VV017277.D
 Acq On : 01 Jul 2020 09:52
 Operator : SY/MD
 Sample : VV0701WBL01
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK58

Quant Time: Jul 02 03:15:30 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR061720WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Jul 02 03:02:49 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	175963	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	171629	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	76208	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	45914	4.35	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	87.00%
7) Chloroethane-d5	1.58	69	41657	5.43	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	108.60%
11) 1,1-Dichloroethene-d2	2.12	63	71866	3.58	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	71.60%
20) 2-Butanone-d5	3.96	46	112830	41.92	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	83.84%
24) Chloroform-d	4.37	84	116522	5.16	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.20%
26) 1,2-Dichloroethane-d4	5.06	65	62597	5.23	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.60%
32) Benzene-d6	5.07	84	205486	4.62	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.40%
36) 1,2-Dichloropropane-d6	6.09	67	61870	4.64	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.80%
41) Toluene-d8	7.33	98	179575	4.30	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.00%
45) trans-1,3-Dichloropropene-	7.64	79	23794	4.29	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.80%
48) 2-Hexanone-d5	8.11	63	85472	40.41	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	80.82%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	42966	4.85	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.00%
65) 1,2-Dichlorobenzene-d4	11.65	152	71478	5.07	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.40%

Target Compounds					Ovalue
16) Methylene chloride	2.52	84	6126	0.677 ug/L	91
49) Tetrachloroethene	8.00	164	2022	0.191 ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV070120\
Data File : VV017277.D
Acq On : 01 Jul 2020 09:52
Operator : SY/MD
Sample : VV0701WBL01
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK58

Quant Time: Jul 02 03:15:30 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR061720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Jul 02 03:02:49 2020
Response via : Initial Calibration

