

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV070120\
 Data File : VV017279.D
 Acq On : 01 Jul 2020 10:47
 Operator : SY/MD
 Sample : L3150-15MS
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 ETYJ9MS

Quant Time: Jul 02 03:31:27 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR061720WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Jul 02 03:02:49 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	183108	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	177510	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	80996	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	46196	4.20	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	84.00%
7) Chloroethane-d5	1.58	69	41743	5.22	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	104.40%
11) 1,1-Dichloroethene-d2	2.13	63	100357	4.80	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	96.00%
20) 2-Butanone-d5	3.93	46	134340	47.96	ug/L	-0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	95.92%
24) Chloroform-d	4.38	84	114686	4.88	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.60%
26) 1,2-Dichloroethane-d4	5.06	65	63815	5.12	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.40%
32) Benzene-d6	5.08	84	202808	4.41	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.20%
36) 1,2-Dichloropropane-d6	6.09	67	61971	4.49	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	89.80%
41) Toluene-d8	7.34	98	179552	4.16	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.20%
45) trans-1,3-Dichloropropene-	7.64	79	23487	4.09	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.80%
48) 2-Hexanone-d5	8.11	63	94459	43.17	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	86.34%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	44302	4.84	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.80%
65) 1,2-Dichlorobenzene-d4	11.65	152	73590	4.92	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.40%

Target Compounds

					Ovalue
12) 1,1-Dichloroethene	2.14	96	47731	5.431 ug/L	94
18) trans-1,2-Dichloroethene	2.78	96	1077	0.116 ug/L	93
22) cis-1,2-Dichloroethene	3.94	96	59833	4.698 ug/L	95
33) Benzene	5.13	78	234683	5.048 ug/L	100
34) Trichloroethene	5.94	95	348989	25.638 ug/L	96
42) Toluene	7.41	91	257562	4.766 ug/L	99
49) Tetrachloroethene	8.00	164	1908	0.174 ug/L	92
53) Chlorobenzene	8.90	112	163383	4.849 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV070120\
Data File : VV017279.D
Acq On : 01 Jul 2020 10:47
Operator : SY/MD
Sample : L3150-15MS
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
ETYJ9MS

Quant Time: Jul 02 03:31:27 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR061720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Jul 02 03:02:49 2020
Response via : Initial Calibration

