

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV070120\
 Data File : VV017289.D
 Acq On : 01 Jul 2020 14:46
 Operator : SY/MD
 Sample : L3150-09
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 ETYJ4

Quant Time: Jul 02 05:32:50 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR061720WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Jul 02 03:02:49 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	175730	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	173150	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	83096	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	40938	3.88	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	77.60%
7) Chloroethane-d5	1.58	69	36775	4.80	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	96.00%
11) 1,1-Dichloroethene-d2	2.12	63	66956	3.34	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	66.80%
20) 2-Butanone-d5	3.96	46	112484	41.85	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	83.70%
24) Chloroform-d	4.38	84	114531	5.08	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.60%
26) 1,2-Dichloroethane-d4	5.06	65	63288	5.29	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.80%
32) Benzene-d6	5.07	84	201692	4.49	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.80%
36) 1,2-Dichloropropane-d6	6.10	67	60643	4.51	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	90.20%
41) Toluene-d8	7.34	98	174145	4.13	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.60%
45) trans-1,3-Dichloropropene-	7.65	79	22347	3.99	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	79.80%
48) 2-Hexanone-d5	8.12	63	92170	43.19	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	86.38%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	45435	5.08	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	101.60%
65) 1,2-Dichlorobenzene-d4	11.65	152	75184	4.89	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.80%

Target Compounds

					Ovalue
22) cis-1,2-Dichloroethene	3.94	96	15575	1.274 ug/L	98
29) 1,1,1-Trichloroethane	4.64	97	4788	0.231 ug/L	91
34) Trichloroethene	5.94	95	441921	33.283 ug/L	98
49) Tetrachloroethene	8.00	164	1686	0.158 ug/L	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV070120\
Data File : VV017289.D
Acq On : 01 Jul 2020 14:46
Operator : SY/MD
Sample : L3150-09
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
ETYJ4

Quant Time: Jul 02 05:32:50 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR061720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Jul 02 03:02:49 2020
Response via : Initial Calibration

