

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW070122\
 Data File : VW026624.D
 Acq On : 01 Jul 2022 13:03
 Operator : SY/MD
 Sample : N3506-03
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COAM2

Quant Time: Jul 04 02:03:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR062322WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Jul 04 02:01:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.622	114	206442	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	194196	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	82421	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.320	65	59774	3.600	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	72.000%	
7) Chloroethane-d5	1.581	69	55241	4.142	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	82.800%	
11) 1,1-Dichloroethene-d2	2.117	63	85787	2.908	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	58.200%#	
20) 2-Butanone-d5	3.915	46	98917	59.928	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	119.860%	
24) Chloroform-d	4.359	84	134310	4.949	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	99.000%	
26) 1,2-Dichloroethane-d4	5.040	65	52033	4.978	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	99.600%	
32) Benzene-d6	5.056	84	230431	4.131	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	82.600%	
36) 1,2-Dichloropropane-d6	6.072	67	73858	4.815	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	96.400%	
41) Toluene-d8	7.317	98	181999	3.628	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	72.600%	
43) trans-1,3-Dichloroprop...	7.625	79	18562	4.682	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	93.600%	
46) 2-Hexanone-d5	8.092	63	72553	48.043	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	96.080%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	50510	5.034	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	100.600%	
66) 1,2-Dichlorobenzene-d4	11.622	152	65350	4.791	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	95.800%	
Target Compounds						
					Qvalue	
14) Carbon disulfide	2.307	76	6232	0.169	ug/L	95
16) Methylene chloride	2.523	84	9636	0.651	ug/L	96
25) Chloroform	4.381	83	197047	7.211	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV070122\
Data File : VV026624.D
Acq On : 01 Jul 2022 13:03
Operator : SY/MD
Sample : N3506-03
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AM2

Quant Time: Jul 04 02:03:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR062322WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Jul 04 02:01:03 2022
Response via : Initial Calibration

