




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV070122\
 Data File : VV026634.D
 Acq On : 01 Jul 2022 17:29
 Operator : SY/MD
 Sample : N3506-13
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AM7

Manual IntegrationsAPPROVED

Quant Time: Jul 04 02:06:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR062322WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Jul 04 02:01:03 2022
 Response via : Initial Calibration

Reviewed By :Krupa Patel 07/05/2022
 Supervised By :Mahesh Dadoda 07/05/2022



TIC: VV026634.D\data.ms

(13) Acetone (T)

2.201min (-0.003) 2.47 ug/L

response 2401

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	37.50	73.80
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV070122\
 Data File : VV026634.D
 Acq On : 01 Jul 2022 17:29
 Operator : SY/MD
 Sample : N3506-13
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AM7

Manual IntegrationsAPPROVED

Quant Time: Jul 04 02:06:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR062322WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Jul 04 02:01:03 2022
 Response via : Initial Calibration

Reviewed By :Krupa Patel 07/05/2022
 Supervised By :Mahesh Dadoda 07/05/2022



TIC: VV026634.D\data.ms

(13) Acetone (T)

2.201min (-0.003) 5.95 ug/L m

response 5798

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	37.50	30.56
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV070122\
 Data File : VV026634.D
 Acq On : 01 Jul 2022 17:29
 Operator : SY/MD
 Sample : N3506-13
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AM7

Manual IntegrationsAPPROVED

Quant Time: Jul 04 02:06:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR062322WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Jul 04 02:01:03 2022
 Response via : Initial Calibration

Reviewed By :Krupa Patel 07/05/2022
 Supervised By :Mahesh Dadoda 07/05/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.625	114	192007	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	185733	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	77463	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.320	65	51409	3.329	ug/L	0.00
Spiked Amount 5.000	Range 40 - 130		Recovery =	66.600%		
7) Chloroethane-d5	1.584	69	48195	3.885	ug/L	0.00
Spiked Amount 5.000	Range 65 - 130		Recovery =	77.800%		
11) 1,1-Dichloroethene-d2	2.121	63	73523	2.680	ug/L	0.00
Spiked Amount 5.000	Range 60 - 125		Recovery =	53.600%#		
20) 2-Butanone-d5	3.921	46	97726	63.657	ug/L	0.00
Spiked Amount 50.000	Range 40 - 130		Recovery =	127.320%		
24) Chloroform-d	4.365	84	120800	4.786	ug/L	0.00
Spiked Amount 5.000	Range 70 - 125		Recovery =	95.800%		
26) 1,2-Dichloroethane-d4	5.047	65	48236	4.961	ug/L	0.00
Spiked Amount 5.000	Range 70 - 130		Recovery =	99.200%		
32) Benzene-d6	5.063	84	201370	3.775	ug/L	0.00
Spiked Amount 5.000	Range 70 - 125		Recovery =	75.400%		
36) 1,2-Dichloropropane-d6	6.079	67	65568	4.469	ug/L	0.00
Spiked Amount 5.000	Range 60 - 140		Recovery =	89.400%		
41) Toluene-d8	7.320	98	158380	3.301	ug/L	0.00
Spiked Amount 5.000	Range 70 - 130		Recovery =	66.000%#		
43) trans-1,3-Dichloroprop...	7.632	79	16110	4.248	ug/L	0.00
Spiked Amount 5.000	Range 55 - 130		Recovery =	85.000%		
46) 2-Hexanone-d5	8.095	63	69054	47.809	ug/L	0.00
Spiked Amount 50.000	Range 45 - 130		Recovery =	95.620%		
56) 1,1,2,2-Tetrachloroeth...	10.217	84	47554	4.955	ug/L	0.00
Spiked Amount 5.000	Range 65 - 120		Recovery =	99.200%		
66) 1,2-Dichlorobenzene-d4	11.625	152	58751	4.583	ug/L	0.00
Spiked Amount 5.000	Range 80 - 120		Recovery =	91.600%		
Target Compounds						
					Qvalue	
13) Acetone	2.201	43	5798m	5.953	ug/L	
14) Carbon disulfide	2.307	76	19642	0.574	ug/L	100
16) Methylene chloride	2.523	84	6128	0.445	ug/L	93
25) Chloroform	4.391	83	133257	5.243	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV070122\
Data File : VV026634.D
Acq On : 01 Jul 2022 17:29
Operator : SY/MD
Sample : N3506-13
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AM7

Manual IntegrationsAPPROVED

Quant Time: Jul 04 02:06:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR062322WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Jul 04 02:01:03 2022
Response via : Initial Calibration

Reviewed By :Krupa Patel 07/05/2022
Supervised By :Mahesh Dadoda 07/05/2022

