

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV070218\
 Data File : VV006532.D
 Acq On : 02 Jul 2018 16:07
 Operator : SY/MD
 Sample : J3792-10
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DB0H5

Quant Time: Jul 02 23:48:56 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR062518WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Jun 30 02:15:06 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	289646	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	265494	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	124908	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	95735	5.03	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	100.60%
7) Chloroethane-d5	1.58	69	84451	5.50	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	110.00%
11) 1,1-Dichloroethene-d2	2.13	63	129281	3.83	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	76.60%
20) 2-Butanone-d5	3.97	46	223559	62.64	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	125.28%
24) Chloroform-d	4.41	84	172908	5.33	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.60%
26) 1,2-Dichloroethane-d4	5.09	65	86562	5.70	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	114.00%
32) Benzene-d6	5.11	84	374375	5.42	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	108.40%
36) 1,2-Dichloropropane-d6	6.12	67	115897	5.63	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	112.60%
41) Toluene-d8	7.37	98	328593	5.01	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.20%
43) trans-1,3-Dichloropropene-	7.67	79	34370	4.53	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.60%
46) 2-Hexanone-d5	8.14	63	177157	58.94	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	117.88%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	74179	5.87	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	117.40%
64) 1,2-Dichlorobenzene-d4	11.68	152	128470	5.54	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	110.80%

Target Compounds					Ovalue
33) Benzene	5.15	78	42283	0.54 ug/L	100
51) Chlorobenzene	8.93	112	195793	3.56 ug/L	100
63) 1,4-Dichlorobenzene	11.33	146	254536	6.19 ug/L	98
65) 1,2-Dichlorobenzene	11.70	146	24910	0.65 ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV070218\
Data File : VV006532.D
Acq On : 02 Jul 2018 16:07
Operator : SY/MD
Sample : J3792-10
Misc : 25 mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DB0H5

Quant Time: Jul 02 23:48:56 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR062518WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Jun 30 02:15:06 2018
Response via : Initial Calibration

