

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV070219\
 Data File : VV011809.D
 Acq On : 02 Jul 2019 17:56
 Operator : SY/MD
 Sample : K3608-10
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GAML9

Quant Time: Jul 03 08:45:24 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR070219WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Jul 03 08:28:44 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	180308	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	170855	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	75155	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	66133	5.45	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	109.00%
7) Chloroethane-d5	1.58	69	52323	5.61	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	112.20%
11) 1,1-Dichloroethene-d2	2.13	63	95252	4.01	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	80.20%
20) 2-Butanone-d5	3.96	46	161546	53.91	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	107.82%
24) Chloroform-d	4.40	84	119251	4.96	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.20%
26) 1,2-Dichloroethane-d4	5.08	65	64521	5.16	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.20%
32) Benzene-d6	5.10	84	244295	5.45	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	109.00%
36) 1,2-Dichloropropane-d6	6.12	67	73238	5.18	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.60%
41) Toluene-d8	7.36	98	211629	5.17	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.40%
43) trans-1,3-Dichloropropene-	7.66	79	24250	4.96	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	99.20%
46) 2-Hexanone-d5	8.14	63	124540	55.26	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	110.52%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	49362	4.96	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	99.20%
64) 1,2-Dichlorobenzene-d4	11.67	152	72449	5.46	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	109.20%

Target Compounds

					Ovalue
13) Acetone	2.22	43	7041	3.431 ug/L	86
14) Carbon disulfide	2.32	76	76626	2.612 ug/L	100
34) Trichloroethene	5.96	95	2343	0.177 ug/L	85
42) Toluene	7.43	91	86682	1.613 ug/L	100
47) Tetrachloroethene	8.02	164	1354	0.130 ug/L	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV070219\
Data File : VV011809.D
Acq On : 02 Jul 2019 17:56
Operator : SY/MD
Sample : K3608-10
Misc : 25ML/MSVOA V/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GAML9

Quant Time: Jul 03 08:45:24 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR070219WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Jul 03 08:28:44 2019
Response via : Initial Calibration

