

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV070219\
 Data File : VV011812.D
 Acq On : 02 Jul 2019 19:09
 Operator : SY/MD
 Sample : K3608-13
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GAMM2

Quant Time: Jul 03 08:55:42 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR070219WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Jul 03 08:28:44 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	172917	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	166597	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	74253	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	64032	5.50	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	110.00%
7) Chloroethane-d5	1.58	69	50910	5.69	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	113.80%
11) 1,1-Dichloroethene-d2	2.13	63	93357	4.10	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	82.00%
20) 2-Butanone-d5	3.94	46	148447	51.65	ug/L	-0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.30%
24) Chloroform-d	4.40	84	116738	5.06	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.20%
26) 1,2-Dichloroethane-d4	5.08	65	63930	5.33	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.60%
32) Benzene-d6	5.10	84	239531	5.48	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	109.60%
36) 1,2-Dichloropropane-d6	6.12	67	72399	5.25	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	105.00%
41) Toluene-d8	7.36	98	205167	5.14	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.80%
43) trans-1,3-Dichloropropene-	7.67	79	22655	4.75	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	95.00%
46) 2-Hexanone-d5	8.13	63	117182	53.32	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	106.64%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	47791	4.92	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.40%
64) 1,2-Dichlorobenzene-d4	11.67	152	71688	5.47	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	109.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV070219\
Data File : VV011812.D
Acq On : 02 Jul 2019 19:09
Operator : SY/MD
Sample : K3608-13
Misc : 25ML/MSVOA V/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GAMM2

Quant Time: Jul 03 08:55:42 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR070219WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Jul 03 08:28:44 2019
Response via : Initial Calibration

