

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV070220\
 Data File : VV017317.D
 Acq On : 02 Jul 2020 19:49
 Operator : SY/MD
 Sample : L3154-02DL 20X
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 H4269DL

Quant Time: Jul 03 06:36:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR061720WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jul 03 06:15:26 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	170809	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	163333	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	80254	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	44084	4.30	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	86.00%
7) Chloroethane-d5	1.58	69	34298	4.60	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	92.00%
11) 1,1-Dichloroethene-d2	2.13	63	66864	3.43	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	68.60%
20) 2-Butanone-d5	3.92	46	94296	36.09	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	72.18%
24) Chloroform-d	4.38	84	93441	4.26	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.20%
26) 1,2-Dichloroethane-d4	5.06	65	50867	4.38	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.60%
32) Benzene-d6	5.08	84	175142	4.13	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.60%
36) 1,2-Dichloropropane-d6	6.10	67	50635	3.99	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	79.80%
41) Toluene-d8	7.34	98	157562	3.96	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.20%
45) trans-1,3-Dichloropropene-	7.65	79	19328	3.66	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	73.20%
48) 2-Hexanone-d5	8.11	63	61952	30.77	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	61.54%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	33898	4.02	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	80.40%
65) 1,2-Dichlorobenzene-d4	11.65	152	63170	4.26	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	85.20%

Target Compounds

					Ovalue
5) Vinyl chloride	1.32	62	86777	7.742	ug/L 99
16) Methylene chloride	2.52	84	4912	0.560	ug/L 90
18) trans-1,2-Dichloroethene	2.78	96	19852	2.285	ug/L 98
22) cis-1,2-Dichloroethene	3.94	96	95496	8.039	ug/L 97
34) Trichloroethene	5.94	95	20446	1.632	ug/L 96
42) Toluene	7.42	91	5485	0.110	ug/L 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV070220\
Data File : VV017317.D
Acq On : 02 Jul 2020 19:49
Operator : SY/MD
Sample : L3154-02DL 20X
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
H4269DL

Quant Time: Jul 03 06:36:56 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR061720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jul 03 06:15:26 2020
Response via : Initial Calibration

