

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV070220\
 Data File : VV017328.D
 Acq On : 03 Jul 2020 00:13
 Operator : SY/MD
 Sample : L3104-17
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Jul 03 07:28:27 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR061720WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jul 03 06:15:26 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	163752	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	161121	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	76906	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	40478	4.12	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	82.40%
7) Chloroethane-d5	1.58	69	32761	4.58	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	91.60%
11) 1,1-Dichloroethene-d2	2.12	63	66201	3.54	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	70.80%
20) 2-Butanone-d5	3.95	46	93290	37.24	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	74.48%
24) Chloroform-d	4.38	84	91987	4.38	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.60%
26) 1,2-Dichloroethane-d4	5.06	65	52057	4.67	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.40%
32) Benzene-d6	5.07	84	171425	4.10	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.00%
36) 1,2-Dichloropropane-d6	6.10	67	47266	3.78	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	75.60%
41) Toluene-d8	7.34	98	153327	3.91	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.20%
45) trans-1,3-Dichloropropene-	7.65	79	19681	3.78	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	75.60%
48) 2-Hexanone-d5	8.11	63	63514	31.98	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	63.96%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	33194	3.99	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	79.80%
65) 1,2-Dichlorobenzene-d4	11.65	152	62362	4.39	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	87.80%

Target Compounds					Ovalue
16) Methylene chloride	2.52	84	5458	0.649 ug/L	86
34) Trichloroethene	5.95	95	3050	0.247 ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV070220\
Data File : VV017328.D
Acq On : 03 Jul 2020 00:13
Operator : SY/MD
Sample : L3104-17
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Jul 03 07:28:27 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR061720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jul 03 06:15:26 2020
Response via : Initial Calibration

